

## SPECIFICATION

Serial no.	Equipment	Specification
1. ✓	pH Meter	Range- 0-14.00 (Digital Reading)
2.	DO Meter	DO range: 0-20.00 mg/L. Oxygen range: 0-200%. ATC: 0-50°. MSC: 0-50 ppt. MAC: 0-20.00 ft.
3. ✓	Microscope	Biological Microscope 30° inclined, 360° rotatable, gemel binocular head Inter-pupillary distance 50-75 mm Plan Achromatic objectives 4X/0.10 Coaxial coarse and line adjustment, with Coarse adjustment stop and tightness adjust. Range: 25mm Precision: 0.002mm
4.	Specimen Jars:	Size: 20ml (Best Quality)
5.	Conical Flask:	Size: 250ml and 500 ml (Best Quality)
6.	Beaker	Size: 250ml (Best Quality)
7.	Formalin	Best Quality
8.	Shelf:	Made of particle board or wooden or with glass doors.